

PCT


WORLD INTELLECTUAL PROPERTY ORGANIZATION  
International Bureau



INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification <sup>5</sup> : <b>G01N 27/26</b>		A2	(11) International Publication Number: <b>WO 94/29704</b>
		(43) International Publication Date: 22 December 1994 (22.12.94)	
(21) International Application Number: <b>PCT/US94/05321</b>		(81) Designated States: AU, CA, JP, KR, NZ, European patent (AT, BE, CH, DE, DK, ES, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE).	
(22) International Filing Date: <b>13 May 1994 (13.05.94)</b>		<p>Published</p> <p><i>With declaration under Article 17(2)(a). Without abstract; title not checked by the International Searching Authority.</i></p>	
(30) Priority Data: 08/073,179      8 June 1993 (08.06.93)      US			
(71) Applicant: <b>BOEHRINGER MANNHEIM CORPORATION</b> [US/US]; 9115 Hague Road, Indianapolis, IN 46250 (US).			
(72) Inventors: <b>WHITE, Bradley, E.</b> ; 3712 Langston Drive, Indianapolis, IN 46268 (US). <b>BROWN, Michael, L.</b> ; 5222 White River Street, Greenwood, IN 46143 (US). <b>RITCHIE, Paul, G.</b> ; 7617 Iron Horse Lane, Indianapolis, IN 46256 (US). <b>SVEINIK, Vladimir</b> ; 539 Cedarlake Court, Carmel, IN 46032 (US). <b>PARKS, Robert, Anthony</b> ; 1447 E. Co. Road, 750 N., Springport, IN 47386 (US). <b>WEINERT, Stefan</b> ; 1557 Martha Street, Fortville, IN 46040 (US).			
(74) Agents: <b>GREEN, Clarence, A. et al.</b> ; Perman & Green, 425 Post Road, Fairfield, CT 06430 (US).			

(54) Title: **BIOSENSING METER WITH AMBIENT TEMPERATURE ESTIMATION METHOD AND SYSTEM**



*FOR THE PURPOSES OF INFORMATION ONLY*

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AT	Austria	GB	United Kingdom	MR	Mauritania
AU	Australia	GE	Georgia	MW	Malawi
BB	Barbados	GN	Guinea	NE	Niger
BE	Belgium	GR	Greece	NL	Netherlands
BF	Burkina Faso	HU	Hungary	NO	Norway
BG	Bulgaria	IE	Ireland	NZ	New Zealand
BJ	Benin	IT	Italy	PL	Poland
BR	Brazil	JP	Japan	PT	Portugal
BY	Belarus	KE	Kenya	RO	Romania
CA	Canada	KG	Kyrgyzstan	RU	Russian Federation
CF	Central African Republic	KP	Democratic People's Republic of Korea	SD	Sudan
CG	Congo	KR	Republic of Korea	SE	Sweden
CH	Switzerland	KZ	Kazakhstan	SI	Slovenia
CI	Côte d'Ivoire	LI	Liechtenstein	SK	Slovakia
CM	Cameroon	LK	Sri Lanka	SN	Senegal
CN	China	LU	Luxembourg	TD	Chad
CS	Czechoslovakia	LV	Latvia	TG	Togo
CZ	Czech Republic	MC	Monaco	TJ	Tajikistan
DE	Germany	MD	Republic of Moldova	TT	Trinidad and Tobago
DK	Denmark	MG	Madagascar	UA	Ukraine
ES	Spain	ML	Mali	US	United States of America
FI	Finland	MN	Mongolia	UZ	Uzbekistan
FR	France			VN	Viet Nam
GA	Gabon				

BIOSENSING METER WITH AMBIENT TEMPERATURE  
ESTIMATION METHOD AND SYSTEM

FIELD OF THE INVENTION

5                   This invention relates to biosensing meters for  
determining the presence of an analyte in a  
biological sample, such determination sensitive to  
ambient temperature, and more particularly, to a  
10                  method and system for estimating such ambient  
temperature.

BACKGROUND OF THE INVENTION

15                Biosensing instruments are often used for the  
detection of various analyte levels in blood  
samples (e.g., glucose and cholesterol). Such  
instruments employ disposable sample strips with a  
well or reaction zone for receiving a blood sample.  
20                Analyte readings obtained from such instruments are  
dependent upon the ambient temperature that  
surrounds the sample well or reaction zone.  
Various prior art instruments employ external  
thermal sensors or make an attempt to control the  
25                temperature of the reaction zone. While external  
temperature sensors are capable of rapidly reacting  
to a temperature change, under certain  
circumstances that becomes a detriment rather than  
an attribute. For instance, if a biosensing  
30                instrument is small enough to be held in a user's  
hand, when that instrument is placed on a tabletop,  
a rapid temperature change can occur that will  
render invalid subsequent biochemical readings -  
until the ambient temperature reading has  
35                stabilized. If the biosensing instrument is

5 battery-driven, it becomes impractical to control the temperature at the reaction zone as such action requires too great a power drain from the instruments battery.

10 The prior art includes a number of disclosures of biosensing instruments that employ temperature correction. In US Patent 5,108,564 to Szuminsky et al, a biosensing instrument is disclosed that measures glucose concentrations in blood. The instrument depends upon a reaction wherein glucose, in the presence of an enzyme, catalyzes a reaction of potassium ferricyanide to potassium ferrocyanide. After the reaction has completed, a voltage is applied across a reaction zone and causes a reversal of the reaction, with an accompanying generation of a small, but measurable current. That current is termed the Cottrell current and, in dependence upon the concentration of glucose in the reaction zone, follows a predetermined curve during the reverse reaction. By determining the position of the curve, an indication of glucose concentration can be obtained.

25 European Patent application 047198682 of Tsutsumi et al discloses a blood glucose measurement system that employs disposable sample strips. The Tsutsumi et al system detects the presence of a blood sample by sensing a resistance across a pair of electrodes. It further employs plurality of sample-like strips, each having a specific resistance value which distinguishes it from other strips. Each of those strips has a particular

application, i.e., for use during an adjustment mode of the instrument, during an error compensation mode; during a calibration mode; etc.

5 US Patent application, 07/451,309, filed December 15, 1989 to White and entitled "Biosensing Instrument and Method" and assigned to the same assignee as this application, teaches a biosensing instrument which employs the "Cottrell" curve  
10 relationship to determine glucose concentrations. In the White patent application, a ratio between current samples and times at which the current samples are taken is used to determine whether the current flow through a sample strip's reaction zone  
15 is, in fact, following the Cottrell relationship.

US Patent 4,420,564 to Tsuji et al. describes a blood sugar analyzer that employs a reaction cell having a fixed enzyme membrane sensor and a  
20 measuring electrode. The Tsuji et al. system includes several fail/safe procedures, one to determine that the reaction is taking place within specifically defined temperature limits and a second to determine if a reaction current remains  
25 within a predetermined range.

In the above noted prior art that indicates a need for temperature sensing, temperature values are obtained by temperature sensors and those sensed  
30 values are directly used. Variations in those sensed temperatures can create substantial variation in biochemical readings and cause erroneous outputs. Since such readings are of

vital importance to the user and, if erroneous, may result in the mis-administration of medications, it is vital that erroneous readings be avoided. Thus, such biosensing instruments must include means for  
5 avoiding erroneous readings that result from erroneous ambient temperature inputs.

Accordingly, it is an object of this invention to provide a biosensing instrument with a method and means for providing accurate temperature values so  
10 as to enable proper analyte value indications.

It is another object of this invention to provide a biosensing instrument with a temperature sensor  
15 that is resistant to rapid temperature excursions as result of environmental changes, but still provides accurate ambient temperature values to enable analyte determinations.

20 SUMMARY OF THE INVENTION

A biosensing meter is provided that determines a value of an analyte in a biological sample. The meter employs an algorithm for determining the  
25 analyte value, which value is dependent upon ambient temperature about the biological sample when it is present in a reaction zone. The biosensing meter includes a processor and a temperature sensor. The temperature sensor is  
30 positioned within the meter's structure and thereby exhibits a delayed response to changes in the ambient temperature. The meter performs an ambient temperature estimation method to overcome the

delayed temperature response. The method commences by the meter repetitively and periodically acquiring temperature readings from the temperature sensor when the biosensing meter is both in an on state and in an off state. When the meter is in the on state, the algorithm estimates the ambient temperature by employing at least two most recent temperature readings and extrapolating therefrom to achieve an ambient temperature estimate. Temperature readings are acquired by the meter at first intervals when the meter is in the off state and at second, shorter intervals when the meter is in the on state, the temperature extrapolations only occurring when the meter is in the on state.

#### DESCRIPTION OF THE DRAWINGS

Fig. 1 is a perspective view of a biosensing meter incorporating the invention.

Fig. 2 is a block diagram of circuitry contained within the biosensing meter of Fig. 1.

Fig. 3 is a waveform diagram illustrating both an excitation voltage applied to an excitation electrode of a disposable sample strip used with the meter of Fig. 1 and a resulting sense current determined from a sense electrode on the disposable sample strip.

Fig. 4 indicates a change in ambient temperature and a resulting change in temperature as sensed at

a temperature sensor within the meter Fig. 1.

Fig. 5 is a high level flow diagram illustrating the procedure followed by the method of the invention.

#### DETAILED DESCRIPTION OF THE INVENTION

Referring now to Fig. 1, a biosensing meter 10 includes a liquid crystal display 12, control buttons 14, and a slot 16 for receiving a disposable sample strip 18. Sample strip 18 contains a well 20 (i.e. reaction zone) that encompasses a pair of conductive electrodes 24 and 26. A layer (not shown) of enzymatic reactants overlays electrodes 24 and 26 in well 20 and provides substrate on which an analyte-containing fluid sample may be emplaced.

Disposable sample strip 18 has an opening 28 that exposes the distal ends of electrodes 24 and 26 and renders them available for electrical connection within biosensing meter 10 (electrical connections not shown in Fig. 1).

A temperature sensor 30 (shown in phantom) is positioned within the case of biosensing meter 10 and provides continuing temperature value inputs to a microprocessor contained within biosensing meter 10. The position of temperature sensor 30 within biosensing meter 10 causes it to be insulated from immediate temperature changes that occur in the ambient about the meter's exterior. As a result,



5 temperature sensor 30 will respond to a change in ambient temperature, but will do so in a delayed manner and over a plurality of thermal time constants, the length of the time constant being a function of the isolation of temperature sensor 30 from the ambient.

10 Referring to Fig. 2, a schematic is shown of circuitry within biosensing meter 10, with a disposable sample strip 18, inserted in slot 16. An excitation voltage source 32 provides a variable voltage to a contact 34 that makes connection with electrode 24 when disposable sample strip 18 is in position within meter 10. A contact 36 enables a  
15 current from electrode 26 to be fed to a sense amplifier 38 whose output (a voltage) is, in turn, fed to an analog to digital converter (A/D) 40. Temperature sensor 30 also provides its output to an A/D converter 42. The outputs from A/D  
20 converters 40 and 42 are applied to a bus 44 which provides communications between modules contained within biosensing meter 10. A microprocessor 46, with an allied display unit 12, provides overall control of the operation of biosensing meter 10.  
25 Microprocessor 46 also has a variety of timing functions 50 contained therein whose use will become apparent from the description below.

30 Excitation voltage source 32 receives its commands from microprocessor 46 via bus 44, and response to those commands applies varying levels of excitation potential to electrode 24. A read only memory key 52 is pluggable into biosensing meter 10 and

contains a non-volatile memory that includes constants and other data required to carry out the analyte-determination procedures required of meter 10. ROM key 52 plugs into the upper most portion of meter 10 as shown in Fig. 1. In general, a ROM key 52 will accompany each batch of disposable sample strips 18, and will contain various constants that will enable meter 10 to adjust its measurement parameters to match the specific batch characteristics of disposable sample strips 18.

In this example, it will be assumed that the analyte-containing sample is a drop of blood that is being subjected to a glucose determination. A disposable sample strip for a glucose determination will include, in well 20, the following reactants: an enzyme, an electrolyte, a mediator, film formers, and a buffer. For instance, the enzyme may be glucose oxidase or glucose dehydrogenase; the buffer may be organic or inorganic; the electrolyte may be potassium chloride or sodium chloride; the mediator is preferably potassium ferricyanide and the film formers comprise gelatin and propiofin. (If the test cell is to be employed for a cholesterol concentration determination, the enzyme would preferably be cholesterol oxidases, with or without a cholesterol esterase additive. The buffer would be preferably inorganic and would include an electrolyte such as potassium chloride or sodium chloride. In this case two mediators would be used, i.e. ferricyanide and quinones, and would be placed in the gelatin film as indicated above.)

As the chemistry employed to make such analyte determinations are known in the art, they will not be described in significant detail. Suffice to say that a glucose determination is made by initially  
5       emplacing in well 20, a sample of blood. The glucose within the sample causes a forward reaction of potassium ferricyanide to potassium ferrocyanide. The forward reaction proceeds to completion during an incubation period. A  
10       subsequent application of an excitation voltage to an electrode in disposable sample strip 18 will see the creation of a small current at the opposite electrode that results from a reverse reaction of potassium ferrocyanide back to potassium  
15       ferricyanide. The flow of electrons during the reverse reaction is sensed and measured at a number of points so as to enable determination that the reaction is both following a Cottrell curve and to further determine the level of the Cottrell curve.  
20       That level is indicative of the glucose concentration. Any resultant glucose value, however, must be corrected to take into account the ambient temperature.

25       The excitation potentials supplied by excitation voltage source 32 to electrode 24 are shown by trace 60 in Fig. 3. The resulting sense current, as determined by sense amplifier 38, is shown by trace 62. Initially, excitation voltage source 32  
30       applies a level 64 to electrode 24. When a blood sample is placed in well 20, a current pulse 66 results, indicating to microprocessor 46 to commence an incubation period. At such time, the

10

excitation voltage level 64 is removed from electrode 24 (level 68) to enable a reaction to occur between the drop of blood and the reactants. At the end of the incubation period, excitation voltage source 32 applies a voltage level 70 to electrode 24. In response, sense amplifier 38 detects and measures a number of currents flowing to electrode 26 (as shown by trace 72).

Assuming that the current sensed across well 20 is following the Cottrell relationship, the current values sensed along curve 72 will be displaced either upwardly or downwardly dependent upon the level of glucose present in the blood sample. Microprocessor 46, in conjunction with ROM key 52 employs the measurements taken along curve 72 to determine the position of curve 72 and derives a glucose value therefrom. In order to alter the resultant glucose value so determined, the ambient temperature must be determined and a correction applied. Only after such correction is made, is the glucose value displayed to the user on display 12.

Turning to Figs. 4 and 5, the temperature estimation procedure employed by a biosensing meter 10 will be described. Because it is desired to isolate temperature sensor 30 from rapid thermal changes, temperature sensor 30 is mounted within biosensing meter 10 so as to be somewhat isolated from the ambient temperature. As a result, when an ambient temperature change occurs, thermal sensor 30 will begin to alter its output in accordance

therewith, but will only arrive at the actual ambient temperature after plural thermal time constants have passed. One thermal time constant for biosensing meter 10 is the time required for temperature sensor 30 to traverse approximately 60 percent of the temperature difference between the beginning temperature of thermal sensor 30 and the actual ambient temperature.

In one meter that has been constructed in accordance with this invention, the thermal time constant is approximately ten minutes. In order for temperature sensor 30 to traverse ninety-eight percent of the temperature between the beginning temperature of thermal sensor 30 and ambient, the passage of four time constants or forty minutes is required-if biosensing meter 10 is to wait for such an occurrence.

Clearly, it is not desirable for the user to wait for forty minutes to obtain an accurate glucose reading. As a result, it has been determined that accurate glucose readings can be obtained by estimating the ambient temperature from a few temperature readings from temperature sensor 30. Those readings are obtained over a short period of time, e.g., characteristically less than one minute. In order to achieve such an estimation of temperature, at least two temperature sensings must be available for calculation by microprocessor 46. One such temperature reading will hereafter be referred to as  $T_{old}$  and a later temperature reading will be referred to as  $T_{new}$ .

So as to not require a user to wait for two temperature readings to be obtained, biosensing meter 10 is caused to take temperature readings at periodic intervals even when it is in the off state. In such off state, sufficient power is provided to microprocessor 46 that microprocessor 46 takes a temperature reading from temperature sensor 30 every three minutes. That temperature reading is stored as  $T_{\text{new}}$  while the former  $T_{\text{new}}$  replaces the previous  $T_{\text{old}}$  value stored in RAM within microprocessor 46. Clearly, those skilled in the art will realize that the time of three minutes is not critical and may be varied in accordance with the meter's requirements.

When meter 10 is subsequently powered on by the user, microprocessor 46 then obtains readings from temperature sensor 30 every thirty seconds to obtain  $T_{\text{new}}$  values. The first such  $T_{\text{new}}$  reading after a power on is combined with the  $T_{\text{old}}$  reading already stored in RAM in microprocessor 46. From those two readings, an initial extrapolation is made to determine  $T_{\text{ambient}}$ . Thereafter, new temperature readings  $T_{\text{new}}$  are taken and at each such time, a  $T_{\text{old}}$  value is dropped and is replaced by the previous  $T_{\text{new}}$  reading which becomes the  $T_{\text{old}}$  value. In such a manner, meter 10 is able to provide  $T_{\text{ambient}}$  values very rapidly after power on and proper glucose value modifications can be made in accordance therewith.

Referring to Fig. 4, it is assumed that  $T_{\text{ambient}}$  varies from level 80 to level 82. Also, it is

assumed that meter 10 is in an off state until 10 minutes have passed at which time it is turned on. Until time  $t=10$ , temperature readings are taken every three minutes (during the off state) and are stored. When the meter is turned on, temperature readings begin being taken every thirty seconds and temperature extrapolations made after every such reading. When  $T_{\text{ambient}}$  changes from level 80 to level 82, the output 84 from temperature sensor 30 begins to approach the new ambient temperature (level 82) in an exponential fashion. Output temperature reading values from temperature sensor 30, taken during this time, enable microprocessor 46 to render an estimation of  $T_{\text{ambient}}$  in accordance with the following relationship:

$$T_{\text{ambient}} = T_{\text{new}} + \frac{T_{\text{new}} - T_{\text{old}}}{e^{m\Delta t} - 1} \quad (A)$$

Where:  $m$  = inverse of the meter's thermal time constant (in seconds); and  
 $\Delta t$  = the time in seconds between acquisition of  $T_{\text{new}}$  and  $T_{\text{old}}$ .

Referring to Fig. 5, the ambient temperature estimation procedure employed by microprocessor 46 will be described. As indicated by decision box 100, if meter 10 is off, a determination is made whether three minutes have elapsed since the last temperature reading (decision box 102). If three minutes have not elapsed, the procedure recycles.

Upon the passage of three minutes, a new temperature reading  $T_{new}$  is taken (box 104). The old temperature reading  $T_{old}$  is replaced with the previous temperature reading  $T_{new}$  (box 106). The procedure then recycles to again determine whether meter 10 has been turned on or off.

Once meter 10 is turned on, the elapsed time  $\Delta t$  from the last temperature reading is determined and compared to a threshold time value obtained from memory. If the  $\Delta t$  value exceeds the threshold value (e.g. thirty seconds), a new temperature reading  $T_{new}$  is taken and stored with  $T_{old}$  (box 110). If the elapsed time  $\Delta t$  is less than the threshold time value, the procedure recycles to wait for the expiration of a proper amount of time.

The aforesaid threshold time value for  $\Delta t$  is employed to prevent microprocessor 46 from making extrapolations based upon temperature readings that are taken too close together in time. Such readings could cause a substantial error which would skew the temperature estimate. As can be seen from equation A, as the value of  $\Delta t$  decreases, the value of  $e^{\Delta t}$  approaches unity. As a result, if  $\Delta t$  is too small the  $1/(e^{\Delta t} - 1)$  factor becomes large. This factor can greatly multiply small errors in the measurement of  $T_{new}$  and  $T_{old}$  and result in an erroneous skew in the ultimate  $T_{ambient}$  calculated value.

Once a new temperature reading  $T_{new}$  is taken, a portion of equation A is calculated (box 112) as



follows:

$$X1 = \frac{1}{e^{m\Delta t} - 1} \quad (B)$$

5        Once equation B is calculated, it is determined  
whether the new sensed temperature  $T_{new}$  minus the old  
temperature  $T_{old}$  is less than a temperature threshold  
value  $\Delta T$  (decision box 114). Threshold value  $\Delta T$  is  
10        obtained from ROM key 52. In the case where both  
 $T_{new}$  and  $T_{old}$  are very close, it is possible that the  
estimation algorithm can actually show a  
significant movement in temperature when, in fact,  
the difference could entirely be due to noise.  
Therefore, the difference between  $T_{new}$  and  $T_{old}$  is  
15        compared to a temperature estimation threshold  $\Delta T$   
obtained from ROM key 52. If the temperature  
difference is greater than the  $\Delta T$  threshold value,  
then a full temperature estimation is warranted.  
If, however, the difference between the temperature  
20        samples is less than the  $\Delta T$  value obtained from ROM  
key 52, the temperature samples are assumed to be  
essentially identical and no temperature movement  
is assumed to have been seen during the entire  $\Delta t$   
(indicating that meter 10 is probably at  
25        equilibrium). Therefore, no estimation is needed  
and  $T_{new}$  may be used as the new ambient temperature

If, by contrast, the temperature difference between  
 $T_{new}$  and  $T_{old}$  is equal to or exceeds the  $\Delta T$  threshold  
30        value, equation (C) is calculated as shown in box 118:

$$T_{\text{ambient}} = T_{\text{new}} + [(T_{\text{new}} - T_{\text{old}})X1] \quad (C)$$

5 Equation (C) allows a new  $T_{\text{ambient}}$  estimate to be obtained. The previously stored value of  $T_{\text{ambient}}$  is then updated with the new value (box 120). The new value  $T_{\text{ambient}}$  is then compared with operating limit values obtained from ROM key 52 (decision box 122).  
10 If  $T_{\text{ambient}}$  is not within proper operating limits, the test is aborted (box 124). If  $T_{\text{ambient}}$  is found to be within proper operating limits, then  $T_{\text{ambient}}$  is employed to compensate the glucose reading value (box 126). Such operating limit values are the  
15 limits of the temperature correction algorithm, (e.g., between 18°C and 32°C).

It should be understood that the foregoing description is only illustrative of the invention.  
20 Various alternatives and modifications can be devised by those skilled in the art without departing from the invention. For example, while two A/D converters 40 and 42 are shown in Fig. 2, a single A/D converter for both inputs will operate  
25 equally well. Accordingly, the present invention is intended to embrace all such alternatives, modifications and variances which fall within the scope of the appended claims.

CLAIMS

1. In a biosensing meter for determining a value of an analyte in a biological sample, a determined analyte value dependent upon an ambient temperature about said biological sample, said biosensing meter including a processor and a temperature sensor, said temperature sensor positioned within said meter and thereby exhibiting a delayed response to changes in said ambient temperature, said meter performing a temperature estimation method under control of said processor so as to overcome said delayed response, said method comprising the steps of:

(a). periodically acquiring temperature readings from said temperature sensor when said biosensing meter is both in an ON state and in an OFF state; and

(b) when said biosensing meter is in said ON state, estimating said ambient temperature by employing at least two most recent temperature readings.

2. The method as recited in claim 1 wherein step b estimates said ambient temperature by using said two most recent temperature readings and extrapolating therefrom to determine said estimate of said ambient temperature.

3. The method as recited in claim 2 wherein said temperature readings are acquired at first

intervals when said meter is in said OFF state, and at second, shorter intervals when said meter is in said ON state, said biosensing meter employing a temperature reading acquired when said meter is in said OFF state and a temperature reading when said meter is in said ON state to determine at least one said temperature estimate.

4. The method as recited in claim 1 wherein said estimation step is only performed if a difference between said two most recent temperature readings exceeds a threshold value, a latest temperature reading being employed as said ambient temperature if said threshold value is not exceeded.

5. The method as recited in claim 1 wherein said estimation step extrapolates said two most recent temperature readings  $T_{old}$  and  $T_{new}$  to obtain  $T_{ambient}$  by employing the expression:

$$T_{ambient} = T_{new} + \frac{(T_{new} - T_{old})}{e^{m\Delta t} - 1}$$

Where:  $m$  = inverse of the meter's thermal time constant (in seconds);  
 $\Delta t$  = the time in seconds between acquisition of  $T_{new}$  and  $T_{old}$ .

6. The method as recited in claim 5 wherein if  $T_{old}$  is acquired when said meter is in said OFF state, and said meter is switched to the ON state and a  $T_{new}$  reading is acquired, the method comprising the added steps of:

determining an elapsed time between when  $T_{old}$  and  $T_{new}$  values are acquired;

comparing said determined elapsed time against a time threshold; and

disregarding said  $T_{new}$  reading if said elapsed time is not at least equal to said elapsed time threshold.

7. The method as recited in claim 5, comprising the following added steps of:

finding a temperature difference value between  $T_{new}$  and  $T_{old}$ ;

comparing said temperature difference value with a temperature difference threshold value; and

employing  $T_{new}$  reading as a new  $T_{ambient}$  value if said temperature difference value is less than said temperature difference threshold value, based upon an assumption that said biosensing meter is at a stable temperature.

8. The method as recited in claim 7 wherein said biosensing meter includes a pluggable read only memory chip, and wherein said temperature difference threshold value is acquired from said pluggable read only memory chip.

9. The method as recited in claim 5, wherein said biosensing meter is enabled to operate only if said  $T_{\text{ambient}}$  falls within preestablished operating temperature limits.

FIG. 1.




FIG. 2.







FIG. 3.FIG. 4.



FIG. 5.



## PATENT COOPERATION TREATY

## PCT

## DECLARATION OF NON-ESTABLISHMENT OF INTERNATIONAL SEARCH REPORT

(PCT Article 17(2)(a) and Rule 39)

Applicant's or agent's file reference 358-924215-P	IMPORTANT DECLARATION	Date of mailing (day/month/year) <b>JUL 13 1994</b>
International application No. PCT/US94/05321	International filing date (day/month/year) 13 MAY 1994	(Earliest) Priority Date (day/month/year) 08 JUNE 1993
International Patent Classification (IPC) or both national classification and IPC Please See Continuation Sheet.		
Applicant BOEHRINGER MANNHEIM CORPORATION		

This International Searching Authority hereby declares, according to Article 17(2)(a), that **no international search report will be established** on the international application for the reasons indicated below.

1. ☒ The subject matter of the international application relates to:
  - a. ☐ scientific theories.
  - b. ☒ mathematical theories.
  - c. ☐ plant varieties.
  - d. ☐ animal varieties.
  - e. ☐ essentially biological processes for the production of plants and animals, other than microbiological processes and the products of such processes.
  - f. ☐ schemes, rules or methods of doing business.
  - g. ☐ schemes, rules or methods of performing purely mental acts.
  - h. ☐ schemes, rules or methods of playing games.
  - i. ☐ methods for treatment of the human body by surgery or therapy.
  - j. ☐ methods for treatment of the animal body by surgery or therapy.
  - k. ☐ diagnostic methods practiced on the human or animal body.
  - l. ☐ mere presentations of information.
  - m. ☐ computer programs for which this International Searching Authority is not equipped to search prior art.
2. ☒ The failure of the following parts of the international application to comply with prescribed requirements prevents a meaningful search from being carried out:  
☐ the description      ☒ the claims      ☐ the drawings
3. ☐ The failure of the nucleotide and/or amino acid sequence listing to comply with the prescribed requirements prevents a meaningful search from being carried out:  
☐ it does not comply with the prescribed standard  
☐ it is not in the prescribed machine readable form
4. Further comments:  
The manipulative steps of determining a value of an analyte were not given in the claims as recited. It was not certain as to what should be done with the algorithm to determine the analyte value.

Name and mailing address of the ISA/US  
Commissioner of Patents and Trademarks  
Box PCT  
Washington, D.C. 20231

Authorized officer  
  
JOHN NIEBLING

Facsimile No. (703) 305-3230

Telephone No. (703) 308-0661

Form PCT/ISA/203 (July 1992)\*

**DECLARATION C NON-ESTABLISHMENT OF  
INTERNATIONAL SEARCH REPORT**

Int .tional application No.  
PCT/US94/05321

The International Patent Classification (IPC) or National Classification and IPC are as listed below:

IPC(5): G01N 27/26  
US CL: 204/403, 406

This Page Blank (USPTO)